

SCHEDA TECNICA

MASCHERINA CHIRURGICA 3 STRATI CON ELASTICO



Descrizione:

MASCHERINA CHIRURGICA 3 STRATI CON ELASTICO

DESCRIZIONE ESTESA

- Mascherina chirurgica 3 strati
- Materiale TNT
- Colore – verde o blu

SURGICAL FACE MASK, 3-PLY, EAR LOOP	
PRODUCT SPECIFICATIONS	
1	GENERAL CHARACTERISTICS 3-ply, with malleable nose piece, single-use, hypoallergenic, fiberglass free, latex free, fluid resistant. (Outer layer color of the mask surface preferably be blue or green)
2	TYPE Ear loop
3	MATERIAL Non-Woven Spunbond and Melblown Fabric
4	BACTERIAL FILTRATION EFFICIENCY (%) B.F.E (ASTM F2101, EN14683:2014 OR OTHER EQUIVALENT STD) >=98% (F2101)
5	PARTICLES FILTRATION EFFICIENCY @ 0.1µM (%) P.F.E (ASTM F2299, EN14683:2014 OR OTHER EQUIVALENT STD) >=98% (F2299)
6	AIR PERMEABILITY (mmH₂O/CM²) (ASTM F2100, EN14683:2014 OR OTHER EQUIVALENT STD) Delta P Between 2 To 5 (MIL-M-36954C)
7	FLUID RESISTANCE (ASTM F2100, EN14683:2014 OR OTHER EQUIVALENT STD) Resistance To Penetration Of Synthetic Blood, Minimum 120mmhg (F1862)
8	FLAMMABILITY (ASTM F2100, EN14683:2014 OR OTHER EQUIVALENT STD) Class 1 (16 CFR 1610)
9	CONFORMANCE TO QUALITY SYSTEM STANDARDS The supplier shall comply with the standard set forth in ISO 13485/13488 or EN 46001/46002 or equivalent standards.



Declaration of Conformity

Certificate No.: TMC200318101-S

Applicant/
Address: **Shenzhen Yueda Electrical Equipment Co., Ltd.**
1-3F, Gangxun Industrial Park, Xikeng, Yuanshan Street, Longgang District,
Shenzhen

Manufacturer/
Address: **Shenzhen Yueda Electrical Equipment Co., Ltd.**
1-3F, Gangxun Industrial Park, Xikeng, Yuanshan Street, Longgang District,
Shenzhen

Product Name: **Disposable respirator**

Trade Name: N/A

Model/Item Number : Disposal 3 layers type

Rated: --

Classification: Class I

Date and
Number of Test Report: March 23,2020
TMC200318101-S

EC-directive: PPE Directive (EU) 2016/425

Test Standard: EN149:2001+A1:2009

Conclusion

This Declaration of PPE Compliance has been granted to applicant based on the results of tests, performed by Laboratory of TMC Testing Services (Shenzhen) Co., Ltd. on sample of the above-mentioned product in accordance with the provisions of the relevant specific standards and the PPE Directive (EU) 2016/425. It is possible to use CE marking to demonstrate the compliance with this Directive.

Place and date of issue: Shenzhen, March 23,2020

TMC Testing Services (Shenzhen) Co., Ltd.

1st Floor, Block A1, Zone A, Xinshidai Gongrong Industrial Park, No. 2,
Shihuan Road, Shiyan Street, Baoan District, Shenzhen, China

Tel: +86-755- 86642861

Email:cert@tmc-lab.com

Http://www.tmc-lab.com

